

TEST REPORT

(Electronic version)



No: 200342194

VERIFICATION WEBSITE: www.gttc.net.cn

VERIFICATION CODE: XLBH-4640-24

ISSUE DATE: 2020-11-09

扫码下载报告



APPLICANT: WUHAN YOUFU INTERNATIONAL TRADE CO., LTD
ADDRESS: ROOM3, 18/F, BUILDING S-1, KAILEGUIYUAN, NO.108, ZHUODAOQUAN ROAD, HONGSHAN DISTRICT, WUHAN CITY, HUBEI PROVINCE

INFORMATION CONFIRMED BY APPLICANT:
DISPOSABLE FACE MASK (NON MEDICAL USE)
QUANTITY: 100 PIECES

STANDARD ADOPTED:
T/CNTAC 55-2020, T/CNITA 09104-2020 <Civil sanitary mask>

DATE RECEIVED/DATE TEST STARTED: 2020-11-02

CONCLUSION:

BACTERIAL FILTRATION EFFICIENCY	M
PARTICLE FILTRATION EFFICIENCY	M
AIRFLOW RESISTANCE	M
BREAKING STRENGTH OF JUNCTION BETWEEN MASK STRING AND MASK BODY	M

NOTE: "M" -MEET THE STANDARD'S REQUIREMENT "F" -FAIL TO MEET THE STANDARD'S REQUIREMENT
"----" -NO COMMENT

REMARK:

THIS REPORT IS THE ENGLISH TRANSLATION VERSION OF THE REPORT 200342193.
THE DECISION INDICATORS ARE DERIVED FROM THE GROUP STANDARD REQUIRED BY CLIENT (T/CNTAC 55-2020, T/CNITA 09104-2020). OUR INSPECTION CAPACITY AUTHORIZED BY CMA COVERS THE INSPECTION ITEMS T/CNTAC 55-2020, T/CNITA 09104-2020.
ALL THE TESTED ITEMS ARE TESTED UNDER THE STANDARD CONDITION (EXCEPT FOR INDICATION).
COPIES OF THE REPORT ARE VALID ONLY RE-STAMPED.
THE EXPERIMENT WAS CARRIED OUT AT No. 1, ZHUJIANG ROAD, PANYU DISTRICT, GUANGZHOU, GUANGDONG, P. R. CHINA.

APPROVED BY:
Yuan Liu ENGINEER

刘圆



PAGE 1 OF 3

TEST REPORT

(Electronic version)

No: 200342194



TEST REPORT

(Electronic version)

No:200342194

BACTERIAL FILTRATION EFFICIENCY (%)

M

(YY 0469-2011 ANNEX B, TEST BACTERIA: STAPHYLOCOCCUS AUREUS ATCC 6538, TEST AREA: 49cm², FLOW RATE: 28.3L/min, MEAN PARTICLE SIZE: 3.0 μm, RESULT OF THE POSITIVE CONTROL: 1.9×10³ CFU, RESULT OF THE NEGATIVE CONTROL: <1CFU)

		REQUIREMENT
BFE ₁	99.3	≥95
BFE ₂	99.2	(T/CNTAC 55-2020, T/CNITA 09104-2020)
BFE ₃	99.2	

PARTICLE FILTRATION EFFICIENCY (%)

M

(YY 0469-2011 5.6.2, AIR FLOW: 30L/min, AEROSOL: NaCl, AEROSOL CONCENTRATION: 15mg/m³, TEMP: 21.9°C, RH: 38.6%)

		REQUIREMENT
MINIMUM	96.42	≥90
		(T/CNTAC 55-2020, T/CNITA 09104-2020)

AIRFLOW RESISTANCE (Pa)

M

(YY 0469-2011 5.6.2, AIR FLOW: 30L/min, AEROSOL: NaCl, AEROSOL CONCENTRATION: 15mg/m³, TEMP: 21.9°C, RH: 38.6%)

		REQUIREMENT
MAXIMUM	47.9	≤49
		(T/CNTAC 55-2020, T/CNITA 09104-2020)

BREAKING STRENGTH OF JUNCTION BETWEEN MASK STRING AND MASK BODY (N)

M

(GB/T 13773.2-2008)
(GB/T 13773.2-2008 idt ISO 13935-2:1999)

		REQUIREMENT
	10	≥5
		(T/CNTAC 55-2020, T/CNITA 09104-2020)



—End of Report—

PAGE 3 OF 5